



This dimension represents 2.0% of 40"SID.
This is the maximum error as allowed By
FDA Specifications.

Step 1:
Place this target on top of the table
Using the turn knobs, size field to
10in x 10in @ 40in SID

Step 2:
Move the target so the lasers align with
the center lines of this target.

Reference Line For Use
During Transverse Light
Field Alignment